



Neo-Natal T4 (Thyroxine)
Total Reaction Time = 2.5hr
Product Code: 2625-300

Intended Use: The Quantitative Determination of Total Thyroxine Concentration in Human (Neonates) whole blood by a Microwell Enzyme Immunoassay

SUMMARY AND EXPLANATION OF THE TEST


Determination of hypothyroidism within the first few days of birth has been recognized as the single most important diagnostic test in neonates by the American Thyroid Association. The need for its early detection and treatment has resulted in the establishment of screening centers by federal and state health departments. A program of early screening of neonates for congenital hypothyroidism was started in Quebec, Canada in the early seventies. They used dry blood spots on filter paper as the sampling device. Very soon the program was followed by other major public health institutions in Canada and the US. By 1978, almost one million infants had been screened and an incidence rate of congenital hypothyroidism was established to be approximately 1 in 7000 births.

Congenital hypothyroidism is probably the single most common preventable cause of mental retardation. Diagnosis and treatment of congenital hypothyroidism within the first 1-2 months after birth appears to be necessary in order to prevent severe mental retardation.

This microplate enzyme immunoassay methodology provides the technician with optimum sensitivity while requiring few technical manipulations. In this method, calibrators, patient specimen, or controls, all made and dried in whole blood are first added to a microplate well. A buffer containing essential ingredients to isolate T4 from blood proteins is added. The blood from the filter paper dots is allowed to elute in the buffer. In the process T4 (Thyroxine) dissociates from the serum (blood) proteins and binds to the antibody that is immobilized on the inside of the microwells. Excess blood is removed using a wash step. Enzyme-T4 conjugate is added. The enzyme labeled T4 binds to the sites on the antibody left available by the native T4 that came from the sample. After the completion of the required incubation period, excess enzyme conjugate is removed using a wash step. The activity of the enzyme present on the surface of the well is quantitated by reaction with a suitable substrate to produce color. The employment of several references, made in whole blood, of known thyroxine concentration permits construction of a dose response curve (DRC-graph) of activity and concentration. An unknown specimen's activity can be extrapolated from the DRC.

PRINCIPLE (TYPE 6):

The essential reagents required for a solid phase sequential enzyme immunoassay include immobilized antibody, enzyme-antigen conjugate and native antigen. Upon mixing immobilized antibody, and a whole blood sample containing the native antigen, a binding reaction results between the native antigen for a limited number of insolubilized binding sites. The interaction is illustrated by the following equation:



Ab_{CW} = Monospecific Immobilized Antibody (Constant Quantity)

Ag = Native Antigen (Variable Quantity)


$AgAb_{CW}$ = Antigen-Antibody Complex

k_a = Rate Constant of Association

k_a = Rate Constant of Disassociation

$K = k_a / k_a$ = Equilibrium Constant

After removing any unreacted native antigen by a wash step, the enzyme-conjugated antigen is introduced. The conjugate reacts with sites of the antibody unoccupied by the native antigen.






$Enz Ag$ = Enzyme-antigen Conjugate (Constant Quantity)



$Enz AgAb_{CW}$ = Enzyme-antigen Conjugate -Antibody Complex

After a short second incubation, the antibody-bound fraction is separated from unbound antigen by decantation or aspiration. The enzyme activity in the antibody-bound fraction is inversely proportional to the native antigen concentration. By utilizing several different calibrators of known antigen concentration, a dose response curve can be generated from which the antigen concentration of an unknown can be ascertained

REAGENTS

Materials Provided:

- A. **Neo-T4 Calibrators – Dried Blood Spots (Two rows by six dots levels - 2 x 6)**
Six (6) levels of T4 Antigen in dried blood spots at approximate concentrations of 0(A), 2 (B), 4(C), 7(D), 14(E) and 28(F) µg/dl placed on S&S type 903 filter paper. Store at 2-8°C. A preservative has been added.
Note 1: The exact values are printed on the outside of the aluminum pouch.
Note 2: The **Lot Specific** calibrators, whole human blood based, were calibrated using analytically pure T4 (greater than 99% by weight). This material exceeds the specifications set by USP.
- B. **Whole Blood Controls – (I, II & III)**
Three (3) controls for thyroxine at varying concentrations (batch specific) made in whole blood spotted on S&S filter paper (Cat# 903) supplied in a zip-lock foil bag with a desiccant. Please see the bag label for ranges for different controls. Store at 2-8°C. A preservative has been added.
- C. **Neo-T4 Elution Buffer:**
One (1) vial containing 13 ml of buffer with binding protein inhibitors, surfactants and preservatives. Store at 2-8°C.
- D. **Neo-T4 Enzyme Diluent – 13 ml –Icon** 
One (1) vial containing 13 ml of buffer, red dye, surfactants and preservatives. Store at 2-8°C.
- E. **NT4 Enzyme Reagent – 1.5ml/vial- Icon** 
One (1) vial of thyroxine-horseradish peroxidase (HRP) conjugate in a protein-stabilizing matrix. A preservative has been added. Store at 2-8°C.
- F. **tT4 Antibody Coated Plate – 96 wells- Icon** 
One 96-well microplate coated with purified Mouse anti-Thyroxine IgG and packaged in an aluminum bag with a drying agent. Store at 2-8°C.

- G. **Wash Solution – 20ml – Icon** 
One (1) vial containing surfactant, buffer and saline. Store at 2-30°C.
- H. **Substrate Solution –12ml/vial - Icon S^N**
One (1) bottle containing tetramethylbenzidine (TMB) and hydrogen peroxide (H₂O₂) in buffer. Store at 2-8°C.
- I. **Stop Solution – 8ml/vial - Icon** 
One (1) bottle containing a strong acid (0.4N HCl). Store at 2-30°C.
- J. **Product Instructions.**

Note 1: Do not use reagents beyond the kit expiration date.

Note 2: Opened reagents are stable for sixty (60) days when stored at 2-8°C.

Note3: Do not use reagents that look cloudy or turbid. They may be contaminated.

Note4: Do not exchange reagents between different batches.

Required But Not Provided:

1. Laboratory Shaker capable of 150rpm rotation.
2. Dispenser(s) for repetitive deliveries of, 0.050ml, 0.100ml and 0.350ml volumes with a precision of better than 1.5%.
3. Adjustable volume (20-200µl) and (200-1000µl) dispenser(s) for conjugate dilutions.
4. 1/8" inch hole punch.
5. Microplate washer or a squeeze bottle (optional).
6. Microplate Reader capable of absorbance readings at 450nm and 620nm.
7. Test tubes for making working enzyme conjugate.
8. Absorbent Paper for blotting the microplate wells.
9. Plastic wrap and microplate cover for incubation steps.
10. Vacuum aspirator (optional) for wash steps.
11. Timer.
12. External quality control materials.

PRECAUTIONS

For In Vitro Diagnostic Use

Not for Internal or External Use in Humans or Animals

All products that contain human blood have been found to be non-reactive for Hepatitis B Surface Antigen, HIV 1&2 and HCV Antibodies by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

SPECIMEN COLLECTION AND PREPARATION

The sampling from neonates is performed by lancing the heels of the infants and then spotting enough whole blood on S&S filter paper card (Cat# 903) to fill the marked circle. Allow the filter paper to dry at room temperature overnight away from heat and moisture. Enclose the dry blood specimen (DBS) in a moisture barrier plastic bag with desiccant and send to the laboratory.

The specimen should be collected 3-7 days post partum, Physical data including age and weight of the infant, whether a multiple birth, or a premature birth etc should accompany the sample. It is important for the clinician to know these facts in order to properly assess the thyroid status of the infant.

The dried blood samples are stable at 2-8°C for 2-3 weeks if stored in zip-lock, moisture resistant bags with desiccants.

REAGENT PREPARATION:

1. **Working T4-Enzyme Conjugate Solution**
Dilute the T4-enzyme conjugate 1:11 with Neo T4 Enzyme Conjugate Diluent in a suitable, clean container. For example, dilute 160µl of conjugate with 1.6ml of buffer for 16 wells (A slight excess of solution is made). **This reagent should be used within two-three hours for maximum performance of the assay.**
General Formula:
Amount of Buffer required = Number of wells * 0.1
Quantity of T4 Enzyme necessary = # of wells * 0.01
i.e. = 16 x 0.1 = 1.6ml for Total T3/T4 Conjugate Buffer
16 x 0.01 = 0.16ml (160µl) for T4 enzyme conjugate
2. **Wash Buffer**
Dilute contents of wash solution to 1000ml with distilled or deionized water in a suitable storage container. Store at 2-30°C.

TEST PROCEDURE

Before proceeding with the assay, bring all reagents and patient samples to room temperature (20 - 27°C).

1. Assemble the required number of microwells for each calibrator, control and patient sample to be assayed in duplicate. **Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.**
2. Punch out 1/8" blood dot out of each calibrator, control and specimens into the assigned wells. **(NOTE: Do not punch blood dots from areas that are printed or that are near the edge of the blood spot).**
3. Add 0.100 ml (100µl) of Neo-T4 Elution Buffer to all the wells.
4. Shake the microplate gently for 20-30 seconds to mix. **(NOTE: Make sure that all blood dots are fully submerged in the liquid and not stuck to the walls of the microwells).**
5. Cover with a microplate cover and rotate for 90 minutes at ambient temperature using a laboratory rotator set @ 150rpm. **(Note: see alternative overnight incubation).**
6. Discard the contents of the microplate by decantation or aspiration. If decanting, blot the plate dry with absorbent paper. **NOTE: Make sure all the blood dots are removed at this point. There should be no dots left in the microwells.**
7. Add 350µl of wash buffer (see Reagent Preparation Section), decant (tap and blot) or aspirate. Repeat four (4) additional times for a total of five (5) washes. **An automatic or manual plate washer can be used. Follow the manufacturer's instruction for proper usage. If a squeeze bottle is employed, fill each well by depressing the container (avoiding air bubbles) to dispense the wash. Decant the wash and repeat four (4) additional times.**
8. Add 100 µl of working Neo-T4 Enzyme Reagent to each well.
9. Cover the microplate and rotate for 45 minutes at ambient temperature using a laboratory rotator set @ 150rpm. **(Note: see alternative overnight incubation).**
10. Repeat wash step #7.
11. Add 0.100 ml (100µl) of substrate solution to each well.
12. Cover the microplate and incubate for 15 minutes at ambient temperature. **No rotation is required for this step.**
13. Add 0.050ml (50µl) of stop solution to each well and gently mix for 15-20 seconds.
NOTE: Always add reagents in the same order to minimize reaction time differences between wells.
14. Read the absorbance in each well at 450nm (using a reference wavelength of 620-630nm to minimize well imperfections) in a microplate reader. **The results should be read within thirty (30) minutes of adding the stop solution.**

Alternative overnight procedure:

1. Substitute overnight incubation (12-16hrs) for the 90 minutes with rotation (Step 5). No rotator is required. Seal the plate(s) with plastic wrap.
2. Substitute 1 hr incubation for the 45 minute incubation with rotation (Step 8). No rotator is required.
3. All other steps remain the same.

CALCULATION OF RESULTS

A dose response curve is used to ascertain the concentration of Neo-natal T4 in unknown specimens.

- Record the absorbance obtained from the printout of the microplate reader as outlined in Example 1.
- Plot the absorbance for each duplicate serum reference versus the corresponding NNT4 concentration in µg/dl on linear graph paper (do not average the duplicates of the serum references before plotting).
- Draw the best-fit curve through the plotted points.
- To determine the concentration of NNT4 for an unknown, locate the average absorbance of the duplicates for each unknown on the vertical axis of the graph, find the intersecting point on the curve, and read the concentration (in µg/dl) from the horizontal axis of the graph (the duplicates of the unknown may be averaged as indicated). In the following example, the average absorbance (0.719) intersects the dose response curve at (10.8µg/dl) NNT4 concentration (See Figure 1).


Note: Computer data reduction software designed for IEMA (ELISA) assays may also be used for the data reduction.

EXAMPLE 1

Sample I.D.	Well Number	Abs (A)	Mean Abs (B)	Value (µg/dl)
Cal A	A1	2.540	2.462	0
	B1	2.492		
Cal B	C1	2.117	2.070	1.4
	D1	2.122		
Cal C	E1	1.761	1.641	3.2
	F1	1.739		
Cal D	G1	1.274	1.094	6.5
	H1	1.247		
Cal E	A2	0.805	0.649	12
	B2	0.824		
Cal F	C2	0.570	0.387	25
	D2	0.576		
Cont - I	E2	1.874	1.855	2.3
	F2	1.836		
Cont - II	G2	1.447	1.436	4.3
	H2	1.425		
Cont - III	A3	0.830	0.785	9.8
	B3	0.740		
Patient	C3	0.698	0.719	10.8
	D3	0.739		

*The data presented in Example 1 and Figure 1 is for illustration only and **should not** be used in lieu of a dose response curve prepared with each assay.

Figure 1



Q. C. PARAMETERS:

In order for the assay results to be considered valid the following criteria should be met.

- The absorbance (OD) of Calibrator '0' µg/dl should be ≥ 1.3 .
- Four out of six quality control pools should be within the established ranges.

QUALITY CONTROL

Each laboratory should assay external controls at levels in the hypothyroid, euthyroid and hyperthyroid range for monitoring assay performance. These controls should be treated as unknowns and values determined in every test procedure performed. Quality control charts should be maintained to follow the performance of the supplied reagents. Pertinent statistical methods should be employed to ascertain trends. The individual laboratory should set acceptable assay performance limits. In addition, maximum absorbance should be consistent with past experience. Significant deviation from established performance can indicate unnoticed change in experimental conditions or degradation of kit reagents. Fresh reagents should be used to determine the reason for the variations

LIMITATIONS OF PROCEDURE

A. Assay Performance

- It is important that the time of reaction in each well is held constant for reproducible results. Pipetting of samples should not extend beyond ten (10) minutes to avoid assay drift. If more than one (1) plate is used, it is recommended to repeat the dose response curve.
- It is very important that blood dots are completely removed from the wells during the first wash step.
- Addition of the substrate solution initiates a kinetic reaction, which is terminated by the addition of the stop solution. Therefore, the addition of the substrate and the stopping solution should be in the same sequence to eliminate any time deviation during reaction.
- Plate readers measure vertically. Do not touch the bottom of the wells.
- Failure to remove adhering solution adequately in the aspiration or decantation wash step(s) may result in poor replication and inaccurate results.
- Microbiologically contaminated samples should not be used in the assay.

EXPECTED RANGES OF VALUES

Based on the limited number of samples at Monobind Inc., and as suggested in the printed literature the normal range for healthy neonates is assigned at 8 – 23 µg/dl.

It is important to keep in mind that any normal range establishment is dependent upon a multiplicity of factors like the specificity of the method, the locale, the population tested and the precision of the method in the hands of technicians. For these reasons each laboratory should depend upon the range of expected values established by the Manufacturer only until an in-house range can be determined by the technicians using the method with a population indigenous to the area in which the laboratory is located.

PERFORMANCE CHARACTERISTICS:

Precision:

The within and between assay precisions of the NNT4 Accubind™ ELISA test system were determined by analyses on three different levels of dried blood controls. The number (N), mean values (X), standard deviation (σ) and coefficient of variation (C.V.) for each of these controls are presented in Table 2 and Table 3.

TABLE 2

Within Assay Precision (Values in µg/dl)

Sample	N	X	σ	C.V.
Low	20	2.76	0.30	10.9%
Normal	20	5.15	0.45	8.8%
High	20	11.30	0.88	7.8%

TABLE 3

Between Assay Precision (Values in µg/dl)

Sample	N	X	σ	C.V.
Low	10	2.86	0.24	8.4%
Normal	10	5.24	0.35	6.7%
High	10	11.10	0.88	7.9%

*As measured in ten experiments in duplicate over a ten day period.

B. Accuracy

The NNT4 Accubind™ ELISA test system was compared with an automated fluorescent methodology. Biological specimens from hypothyroid, euthyroid and hyperthyroid populations were used (The values ranged from 0.5 µg/dl – 46 µg/dl). The total number of such specimens was 370. The least square regression equation and the correlation coefficient were computed for this NNT4 Accubind™ ELISA method in comparison with the reference method. The data obtained is displayed in Table 4.

TABLE 4
Least Square
Regression
Analysis

Method	Mean	Correlation Coefficient
This Method (y)	15.63	0.955
Reference (x)	15.96	

Only slight amounts of bias between this method and the reference method are indicated by the closeness of the mean values. The least square regression equation and correlation coefficient indicates excellent method agreement.

C. Sensitivity

The NNT4 Accubind™ ELISA test system has a sensitivity of 0.5 µg/dl. The sensitivity was ascertained by determining the variability of the 0 µg/dl calibrator and using the 2 σ (95% certainty) statistic to calculate the minimum dose.

REFERENCES

- Harrison, *Internal Medicine*, Mosby Printing, Philadelphia, 12th Ed.
- Dussault JH, Coloumbe C, Laberge C, Latarte J, Guyde H and Khoury K, "Preliminary report on Mass Screening Program for neonatal Hypothyroidism", *J. Pediatr*, **86**, 670-674 (1975).
- Dussault JH & Laberge C, "Thyroxine (T4) determination in dried blood by radioimmunoassay: A screening Method for Neonatal Hypothyroidism", *Union Med Con*, **102**, 2062-64 (1973).
- Fisher DA, "Neonatal Thyroid Screening", *Pediatric Clinics of North America*, **25**, 423-430 (1978).
- Klein, A., and T.P. Foley, *Lancet*, **2**, 77 (1974).
- Biosafety in Biomedical and Microbiological laboratories*, 4th Ed, U.S. Department of Health and Human Services, HHS Publication No 93-8395 (CDC), Washington DC, USA (1999).
- Young DS, Pestaner LC, & Gibberman V, *Clinical Chemistry*, **21**, 5 (1975).

Revision: D

Date: 011707

Cat# 2625-300			
Size	96(A)	192(B)	
Reagent (fill)	A)	1 Card	1 Card
	B)	1 Card	1 Card
	C)	1 (13ml)	2 (13ml)
	D)	1 (13ml)	2 (13ml)
	E)	1 (1.5ml)	2 (1.5ml)
	F)	1 plate	2 plates
	G)	1 (20ml)	1 (20ml)
	H)	1 (12ml)	2 (12ml)
	I)	1 (8ml)	2 (8ml)

For Orders and Inquiries, please contact



Tel: 949-951-2665
 Fax: 949-951-3539
 Email: info@monobind.com
 On the Web: www.monobind.com

Please visit our website to learn more about our other interesting products and services.



EC REP CEpartner4U, 3951 DB; 13.NL
 Tel: +31 (0) 6-516.536.26